

PURERAY INSTRUMENT CO., LTD

CALIBRATION CERTIFICATE

Certificate No. : PR2026J0513062

Name of Instrument: X-ray Fluorescence Spectrometer

Type/specification: XRF-K5

Manufacturer: PURERAY INSTRUMENT CO., LTD

Serial No.: ZS/N: 6643A83F1903

Reference Documents: JJF1133-2005 《C.S. for X-Ray Fluorescence Gold Content Analyzers》

Caution:

The calibration certificate contains data and information applies only to the calibrated instrument.

Copying of using select parts or tampering with this document without the permission of our company is forbidden. Copies of this certificate without an official stamp of calibration are not valid.

Major standards of measurement used in the calibration and test result:

Standard Sample Name	Standard Value	Substance Number	Test Result	Indication Error	Traceability Due Date
Gold Sample	87.48%	GBW(E)020238	87.56%	+0.08	2036.3
Gold Sample	75.03%	GBW(E)020240	74.97%	-0.06	2036.3
Gold Sample	42.00%	GBW(E)020248	42.02%	+0.02	2036.3

Operation Date : 2026.5.13

This product has passed the calibration inspection.

Approved by: Zllie

Checked by: Aemy

Calibrated by: Howard Ye

Certificate Issuing Organization

